



TÜVRheinland[®]
Precisely Right.

CERTIFICATE

Conformity Assessment for Analysis Equipment

Certificate No.: AK 28181126

Name and address of the Holder : **PCF Elettronica S.r.l.**
Via F.lli Savio, 31
24040 Levate (BG)
ITALY

Type of equipment : **VOC analyzer (portable analyzer)**

Identification : **Model : FID 2001/C**

Description : **Continuous analyser for the measurement of organic compounds**
smallest tested measuring range : 0 ...100 mg / Nm³
limit of detection : <= 0,7 % range

Reference specification : **DLgs Aprile 2006, n. 152 parte V, titoli I e II**

Report No.: **28181126**

We hereby certify, the conformity of the a.m. equipment which complies with the relevant requirements of the reference specification.

During the periodical monitoring, compliance with the specification has been demonstrated. The monitoring will be performed according to the holder's calibration procedure

Milan, 2008-10-30



TÜV CERT-Certification Body for
Pressure Equipment


Dipl.-Ing. J. Moreno
Notified Body, ID-No. 0035

TÜV Rheinland Industrie Service GmbH
Am Grauen Stein
D-51105 Cologne

TÜV Rheinland Italia
I - 24030 Medolago (BG)
Tel. : +39 035 49.33.147
e-mail : bergamo@it.tuv.com

This Certificate does not imply assessment of the equipment manufacturing and does not permit the use of a TÜV Rheinland mark of conformity